

## New England Biolabs Certificate of Analysis

**Product Name:** *Haelll*  
**Catalog Number:** *R0108T*  
**Concentration:** *50,000 U/ml*  
**Unit Definition:** *One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in rCutSmart Buffer in 1 hour at 37°C in a total reaction volume of 50 µl.*  
**Packaging Lot Number:** *10202245*  
**Expiration Date:** *07/2025*  
**Storage Temperature:** *-20°C*  
**Storage Conditions:** *10 mM Tris-HCl, 50 mM KCl, 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml rAlbumin (pH 7.4 @ 25°C)*  
**Specification Version:** *PS-R0108T/M v2.0*


Haelll Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0108TVIAL	Haelll	10200030	Pass
B7024AVIAL	Gel Loading Dye, Purple (6X)	10193044	Pass
B6004SVIAL	rCutSmart™ Buffer	10196037	Pass

Assay Name/Specification	Lot # 10202245
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 100 units of Haelll incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
<b>Functional Testing (15 minute Digest)</b> A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of Lambda DNA and 1 µl of Haelll incubated for 15 minutes at 37°C results in complete digestion as determined by agarose gel electrophoresis.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> After a 20-fold over-digestion of Lambda DNA with Haelll, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Haelll.	Pass
<b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of	Pass

Assay Name/Specification	Lot # 10202245
<p>100 units of HaeIII incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	
<p><b>Protein Purity Assay (SDS-PAGE)</b> HaeIII is &gt;95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.</p>	<b>Pass</b>
<p><b>qPCR DNA Contamination (E. coli Genomic)</b> A minimum of 10 units of HaeIII is screened for the presence of E. coli genomic DNA using SYBR<sup>®</sup> Green qPCR with primers specific for the E. coli 16S rRNA locus. Results are quantified using a standard curve generated from purified E. coli genomic DNA. The measured level of E. coli genomic DNA contamination is <math>\leq 1</math> E. coli genome.</p>	<b>Pass</b>


This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.




---

YunJie Sun  
Production Scientist  
24 Jul 2023




---

Michael Tonello  
Packaging Quality Control Inspector  
10 Aug 2023