

## New England Biolabs Certificate of Analysis

**Product Name:** BsmI  
**Catalog Number:** R0134L  
**Concentration:** 10,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 65°C in a total reaction volume of 50 µl.  
**Lot Number:** 10026223  
**Expiration Date:** 10/2020  
**Storage Temperature:** -20°C  
**Storage Conditions:** 50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA  
**Specification Version:** PS-R0134S/L v1.0

| BsmI Component List |                       |            |                      |
|---------------------|-----------------------|------------|----------------------|
| NEB Part Number     | Component Description | Lot Number | Individual QC Result |
| R0134LVIAL          | BsmI                  | 10026224   | Pass                 |
| B7204SVIAL          | CutSmart® Buffer      | 10018443   | Pass                 |

| Assay Name/Specification   | Lot # 10026223 |
|--|----------------|
| <b>Exonuclease Activity (Radioactivity Release)</b><br>A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 100 units of BsmI incubated for 4 hours at 65°C releases <0.1% of the total radioactivity.   | Pass           |
| <b>Ligation and Recutting (Terminal Integrity)</b><br>After a 10-fold over-digestion of Lambda DNA with BsmI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 4 hours at 25°C. Of these ligated fragments, >95% can be recut with BsmI.   | Pass           |
| <b>Non-Specific DNase Activity (16 Hour)</b><br>A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pBR322 DNA and a minimum of 30 units of BsmI incubated for 16 hours at 65°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. | Pass           |

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso  
Production Scientist  
12 Sep 2018



Michael Tonello  
Packaging Quality Control Inspector  
23 Oct 2018