

New England Biolabs Certificate of Analysis


Product Name: Sall
Catalog Number: R0138M
Concentration: 100,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA (HindIII digest) in 1 hour at 37°C in a total reaction volume of 50 µl.
Lot Number: 10054477
Expiration Date: 09/2021
Storage Temperature: -20°C
Storage Conditions: 50 mM KCl, 10 mM Tris-HCl (pH 7.5), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 300 µg/ml BSA
Specification Version: PS-R0138T/M v1.0

Sall Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0138M VIAL	Sall	10054478	Pass
B7203S VIAL	NEBuffer™ 3.1	10053972	Pass
B7024S VIAL	Gel Loading Dye, Purple (6X)	10050274	Pass


Assay Name/Specification	Lot # 10054477
Blue-White Screening (Terminal Integrity) A sample of pUC19 vector linearized with a 10-fold excess of Sall, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in <1% white colonies.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of Sall incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of Adenovirus-2 DNA with Sall, >95% of the DNA fragments can be ligated with T4 DNA ligase in 4 hours at 25°C. Of these ligated fragments, >95% can be recut with Sall.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of pBR322 DNA and a minimum of 20	Pass

Assay Name/Specification	Lot # 10054477
units of Sall incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	

This product has been tested and shown to be in compliance with all specifications.



Doreen Duquette
Production Scientist
26 Aug 2019



Jay Minichiello
Packaging Quality Control Inspector
30 Sep 2019