

New England Biolabs Certificate of Analysis


Product Name: *AvaI*
Catalog #: *R0152T/M*
Concentration: *50,000 units/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.*
Lot #: *0511412*
Assay Date: *12/2014*
Expiration Date: *12/2016*
Storage Temp: *-20 °C*
Storage Conditions: *50 mM KCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA*
Specification Version: *PS-R0152T/M v1.0*
Effective Date: *15 Jul 2013*

Assay Name/Specification (minimum release criteria)	Lot #0511412
Blue-White Screening (Terminal Integrity) - A sample of pUC19 vector linearized with a 10-fold excess of <i>AvaI</i> , religated and transformed into an <i>E. coli</i> strain expressing the LacZ beta fragment gene results in <1% white colonies.	Pass
Exonuclease Activity (Radioactivity Release) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 100 units of <i>AvaI</i> incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) - After a 20-fold over-digestion of Lambda DNA with <i>AvaI</i> , >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with <i>AvaI</i> .	Pass
Non-Specific DNase Activity (16 Hour) - A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 100 Units of <i>AvaI</i> incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.



Authorized by
Derek Robinson
15 Jul 2013



Inspected by
Terry Petronzio
18 Dec 2014

