240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

New England Biolabs Certificate of Analysis

 Product Name:
 Tth111I

 Catalog #:
 R0185S/L

 Concentration:
 5,000 units/ml

 Lot #:
 0351602

 Assay Date:
 02/2016

 Expiration Date:
 2/2018

 Storage Temp:
 -20°C

Storage Conditions: 500 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA

Specification Version: PS-R0185S/L v1.0
Effective Date: 30 Apr 2014

Assay Name/Specification (minimum release criteria)	Lot #0351602
Exonuclease Activity (Radioactivity Release) - A 50 μl reaction in CutSmart™ Buffer containing 1 μg of a mixture of single and double-stranded [³ H] <i>E. coli</i> DNA and a minimum of 50 units of Tth111I incubated for 4 hours at 65°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) - After a 5-fold over-digestion of pBC4 DNA with Tth111I, ~25% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Tth111I.	Pass
Non-Specific DNase Activity (16 hour) - A 50 μl reaction in CutSmart TM Buffer containing 1 μg of pBC4 DNA and a minimum of 5 units of Tth111I incubated for 16 hours at 65°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme.	Pass
Protein Purity Assay (SDS-PAGE) - Tth1111 is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass

^{*} The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.

Authorized by Derek Robinson 30 Apr 2014







Inspected by
Anthony Francis
18 Feb 2016