

New England Biolabs Certificate of Analysis

Product Name: Tth1111
Catalog Number: R0185S
Concentration: 5,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of pBC4 DNA in 1 hour at 65°C in a total reaction volume of 50 µl.
Lot Number: 10035465
Expiration Date: 02/2021
Storage Temperature: -20°C
Storage Conditions: 500 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0185S/L v1.0

Tth1111 Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0185SVIAL	Tth1111	10035466	Pass
B7204SVIAL	CutSmart® Buffer	10021124	Pass

Assay Name/Specification	Lot # 10035465
<p>Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [³H] E. coli DNA and a minimum of 50 units of Tth1111 incubated for 4 hours at 65°C releases <0.1% of the total radioactivity.</p>	Pass
<p>Ligation and Recutting (Terminal Integrity) After a 5-fold over-digestion of pBC4 DNA with Tth1111, ~25% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Tth1111.</p>	Pass
<p>Non-Specific DNase Activity (16 hour) A 50 µl reaction in CutSmart™ Buffer containing 1 µg of pBC4 DNA and a minimum of 5 units of Tth1111 incubated for 16 hours at 65°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis. NOTE: although no nuclease degradation is detected under these conditions, extended incubations and/or high concentrations of this enzyme may result in star activity. See the product FAQ for recommended reaction conditions for this enzyme.</p>	Pass
<p>Protein Purity Assay (SDS-PAGE)</p>	Pass

Assay Name/Specification	Lot # 10035465
Tth1111 is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	

This product has been tested and shown to be in compliance with all specifications.



Tony Spear-Alfonso
Production Scientist
24 Jan 2019



Michael Tonello
Packaging Quality Control Inspector
21 Feb 2019