

New England Biolabs Certificate of Analysis


Product Name: Acc65I
Catalog Number: R0599S
Concentration: 10,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of pBC4 DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Lot Number: 10044861
Expiration Date: 05/2021
Storage Temperature: -20°C
Storage Conditions: 100 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R0599S/L v1.0

Acc65I Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0599SVIAL	Acc65I	10044862	Pass
B7203SVIAL	NEBuffer™ 3.1	10041637	Pass

Assay Name/Specification	Lot # 10044861
Endonuclease Activity (Nicking) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of supercoiled PhiX174 DNA and a minimum of 50 Units of Acc65I incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Blue-White Screening (Terminal Integrity) A sample of Litmus28i vector linearized with a 10-fold excess of Acc65I, religated and transformed into an E. coli strain expressing the LacZ beta fragment gene results in <1% white colonies.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of Acc65I incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in NEBuffer 3.1 containing 1 µg of pBC4 DNA and a minimum of 100 Units of Acc65I incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass

Assay Name/Specification	Lot # 10044861
<p>Ligation and Recutting (Terminal Integrity) After a 20-fold over-digestion of pBC4 DNA with Acc65I, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Acc65I.</p>	<p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.



Doreen Duquette
Production Scientist
05 Apr 2019



Michael Tonello
Packaging Quality Control Inspector
03 Jul 2019